



2311 West 22nd Street, Suite 200
Oak Brook, Illinois, 60523 • USA
Phone 630-954-1250 • Fax 630-954-1275

January 16, 2020

Certificate of Conformance

We hereby certify that *Taylor Model 8791, Dishwasher Thermometer*, was manufactured, calibrated and inspected to meet design requirements. Accuracy of this instrument, when used in the manner for which it was intended is as follows:

Temperature Range: 32° to 194°F
 0° to 90°C

Accuracy: ±2.0°F for range of 32° - 194° F
 ± 1°C for a range of 0° - 90°C

IP67 & NIST Traceable calibration certified
NSF Listed

Jeff Koon
Director of Engineering