



T&S BRASS AND BRONZE WORKS, INC.

2 Saddleback Cove / P.O. Box 1088
Travelers Rest, SC 29690



Model No.

BL-4203 SERIES

Item No.

Travelers Rest, SC: 800-476-4103 Simi Valley, CA: 800-423-0150 Fax: 864-834-3518 www.tsbrass.com

This Space for Architect/Engineer Approval

Job Name _____ Date _____

Model Specified _____ Quantity _____

Customer/Wholesaler _____

Contractor _____

Architect/Engineer _____


BL-4203-0 Turret w/ (2) Hose Cocks @ 90°

BL-4203-01 Turret w/ (1) Hose Cock

BL-4203-02 Turret w/ (2) Hose Cocks @ 180°


BL-4203-03 Turret w/ (3) Hose Cocks @ 90°

BL-4203-04 Turret w/ (4) Hose Cocks @ 90°



Use BL-4010-01 Gas Cock for Oxygen
or for Pressures above 0.5 PSIG

BL-4203-01 SHOWN



Product Specifications:

**Laboratory Turret w/ Serrated Hose Cock
& Vandal Resistant Pin**

Drawn
ARH

Checked
JRM

Approved
JHB

Scale:

1:2

Date:

8/21/08

Sheet: 1 of 2



T&S BRASS AND BRONZE WORKS, INC.

2 Saddleback Cove / P.O. Box 1088
Travelers Rest, SC 29690



REG. #A2601
ISO #9001

Model No.

BL-4203 SERIES

Item No.

Travelers Rest, SC: 800-476-4103 Simi Valley, CA: 800-423-0150 Fax: 864-834-3518 www.tsbrass.com


ITEM NO.	SALES NO.	DESCRIPTION
1	001203-45	Cap, Index, Orange, Air
2	001200-45	Cap, Index, Yellow, Vacuum
3	001197-45	Cap, Index, Blue, Gas
4	BL-4000-01	Ground Key Hose Cock w/ Blue Gas Cap Index
5	144L	Gasket Polypropylene



① For Air Application




② For Vacuum Application



③ For Gas Application

④



Turret Sold Separately
w/ Gasket as BL-4102-XX.
-XX Suffix References the
Number of Outlets and their
Orientation per Turret. See
Table on Page "1" for Suffix

⑤

Product Specifications:

Laboratory Turret w/ Serrated Hose Cock
& Vandal Resistant Pin

Drawn

ARH

Checked

JRM

Approved

JHB

Scale:

1:2

Date:

8/21/08